Human Phospho-P53 (Ser15) Sandwich ELISA Kit Datasheet

Please read it entirely before use

Catalogue Number: KE40001

Size: 96T

Sensitivity: 2.3 ug/mL

Range: N/A

Usage: For the quantitative detection of Human Phospho-P53 (Ser15) in samples.

This product is for research use only and not for use in human or animal therapeutic or diagnostic.

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

Table of content page

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

1. Background

Stabilized p53 transcriptionally activates genes encoding proteins involved in cell cycle arrest, DNA repair, and/or apoptosis, and represses transcriptional activation of growth-promoting genes. Under physiological conditions, p53, in most cells, is expressed at a low or undetectable level, with a half-life of a few minutes. Under a condition of stress, such as DNA damage caused by ionizing radiation or cytotoxic agents, endogenous p53 is stabilized through a series of physiological responses, including ataxia telangiectasia mutated (ATM)/ATM and Rad3-related (ATR) activation, phosphorylation and acetylation of p53, and weakening of the binding of MDM2 to p53. Phosphorylation of serine 15 also prevents p53 from being exported from the nucleus and stimulates p53 transcriptional activity through the increased association with p300 coactivator.

2. Principle



Sandwich ELISA structure (Detection antibody labeled with biotin)

A capture antibody is pre-coated onto the bottom of wells which binds to analyte of interest. A detection antibody labeled with biotin also binds to the analyte. Streptavidin-HRP binds to the biotin. TMB acts as the HRP substrate and the solution color will change from colorless to blue. A stop solution containing sulfuric acid turns solution yellow. The color intensity is proportional to the quantity of bound protein which is measurable at 450 nm with the correction wavelength set at 630 nm.

3. Required Materials

3.1 A microplate reader capable of measuring absorbance at 450 nm with the correction wavelength set at 630 nm.

3.2 Calibrated, adjustable precision pipettes and disposable plastic tips. A manifold multi-channel pipette is recommended for large assays.

3.3 Plate washer: automated or manual.

3.4 Absorbent paper towels.

3.5 Glass or plastic tubes to prepare standard and sample dilutions.

3.6 Beakers and graduated cylinders.

3.7 Log-log or semi-log graph paper or computer and software for ELISA data analysis. A four-parameter logistic (4-PL) curve-fit is recommended.

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

4. Kit Components and Storage

Microplate - antibody coated 96-well microplate (8 well × 12 strips)	1 plate	**Unopened Kit:** Store at 2-8°C for 6 months or -20°C for 12 months. **Opened Kit:** All reagents stored at 2-8°C for 7 days. **Please use a new standard for each assay.**
Protein standard - 2000 pg/bottle; lyophilized	2 bottles	
Detection Antibody (100×), biotinylated - 120 µL/vial*	1 vial	
Streptavidin-horseradish peroxidase (HRP) (100×) - 120 µL/vial*	1 vial	
Sample Diluent PT 3-ef - 30 mL/bottle. For Human serum and plasma	1 bottle	
Sample Diluent PT 4-ef - 30 mL/bottle. For Mouse/Rat serum, plasma and serum-free cell culture supernatants.	1 bottle	
Sample Diluent PT 5-ef - 30 mL/bottle. For tissue lysates.	1 bottle	
Detection Diluent - 30 mL/bottle	1 bottle	
Wash Buffer Concentrate (20×) - 30 mL/bottle	1 bottle	
Extraction Reagent - 30 mL/bottle	1 bottle	
Tetramethylbenzidine Substrate (TMB) - 12 mL/bottle	1 bottle	
Stop Solution - 12 mL/bottle	1 bottle	
Plate Cover Seals	4 pieces	

* Centrifugation immediately before use

5. Safety Notes

5.1 Avoid any skin and eye contact with Stop Solution and TMB. In case of contact, wash thoroughly with water.

5.2 Do not use the kit after the expiration date.

5.3 Do not mix or substitute reagents or materials from other kit lots or other sources.

5.4 Be sure to wear protective equipment such as gloves, masks and goggles during the experiment.

5.5 When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer to improve assay precision

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

6. Sample Collection and Storage

6.1 Cell Lysate:

1) Collect cells and wash by centrifuging at 500 x g for 5 minutes before resuspension in pre-cooled PBS buffer. Perform this step three times.

2) Count cells and then discard the supernatant.

3) Add protease inhibitor cocktail to the Extraction Reagent to a final concentration immediately prior to performing cell lysis.

4) Add 1 mL of Extraction reagent (containing protease inhibitor cocktail) Per 1 x 107 cells, Incubate cell suspension on ice for 30 minutes, use ultrasound to treat the samples.

5) Centrifuge cell lysate at 10,000 x g for 5 minutes at 4°C.

6) Measure the concentration of total protein in cell lysate using BCA assay. Where possible, keep samples on ice to avoid protein degradation.

6.2 Tissue Lysate:

1) Rinse tissue with PBS, cut into 1-2 mm pieces.

2) Add protease inhibitor cocktail to the Extraction Reagent to a final concentration immediately prior to performing tissue lysis.

3) Add 1 mL of Extraction Reagent containing protease inhibitor cocktail per 100 mg tissue.

4) Homogenize the tissue completely using desired method on ice, Incubate on ice for 30 minutes, use ultrasound to break up the cells.

5) Centrifuge tissue homogenates at 10,000 x g for 5 minutes at 4°C. Collect the supernatant, assay immediately or aliquot and store at -20°C.

6) Measure the concentration of total protein in tissue homogenates using BCA assay.

7) Avoid protein degradation by performing all the above procedures on ice where possible.

6.3 Tissue Homogenates: The preparation of tissue homogenates will vary depending upon tissue type. Rinse tissue with 1 x PBS to remove excess blood, Then add 1 mL of 1 x PBS per 100 mg tissue, homogenized and stored overnight at -20°C. After two freeze-thaw cycles were performed to break the cell membranes, the homogenates were centrifuged for 5 minutes at 10000 x g at 4°C. Collect the supernatant, assay immediately or aliquot and store at -20°C. Avoid repeated freeze-thaw cycles.

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

7. Regent Preparation

7.1 Wash Buffer (1X): If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 30 mL of Wash Buffer Concentrate(20X) to 570 mL deionized or distilled water to prepare 1X Wash Buffer.

7.2 Detection Antibody (1X): Dilute 100X Detection Antibody 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 µL 100X Detection Antibody + 990 µL Detection Diluent (Centrifuge the 100 X Detection Antibody solution for a few seconds prior to use).

7.3 Streptavidin-HRP (1X): Dilute 100X Streptavidin-HRP 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 µL 100X Streptavidin-HRP + 990 µL Detection Diluent (Centrifuge the 100X Streptavidin-HRP solution for a few seconds prior to use).

7.4 Sample Dilution: Different samples should be diluted with corresponding Sample Diluent, samples may require further dilution if the readout values are higher than the highest standard OD reading. Variations in sample collection, processing and storage may affect the results of the measurement.

Recommended Dilution for different sample types: 1:2 is recommended for serum and plasma; 1:2 is recommended for cell culture supernatant .

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

8. Assay Procedure Summary

Bring all reagents to room temperature before use (Detection antibody and Streptavidin-HRP can be used immediately). To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

8.1 Take out the required number of microplate strips and return excess strips to the foil pouch containing the drying reagent pack and reseal; store at 4°C immediately. Microplate strips should be used in one week.

8.2 Preset the layout of the microplate, including control group,standard group and sample group, add 100 µL of each standard and sample to the appropriate wells.(Make sure sample addition is uninterrupted and completed within 5 to 10 minutes, It is recommended to assay all standards, controls, and samples in duplicate).

8.3 Seal plate with cover seal, pressing it firmly onto top of microwells. Incubate the plate for 2 hours at 37°C.

8.4 Wash

1) Gently remove the cover seal. Discard the liquid from wells by aspirating or decanting. Remove any residual solution by tapping the plate a few times on fresh paper towels.

2) Wash 4 times with 1X Wash Buffer, using at least 350-400 µL per well. Following the last wash, firmly tap plates on fresh towels 10 times to remove residual Wash Buffer. Avoid getting any towel fibers in the wells or wells drying out completely.

8.5 Add 100 µL of 1X Detection Antibody solution (refer to Reagent Preparation7.2) to each well. Seal plate with cover seal and incubate for 1 hour at 37°C.

8.6 Repeat wash step in 8.4.

8.7 Add 100 µL of 1X Streptavidin-HRP solution (refer to Reagent Preparation7.3) to each well. Seal plate with cover seal and incubate the plate for 40 minutes at 37°C .

8.8 Repeat wash step in 8.4.

8.9 Signal development: Add 100 µL of TMB substrate solution to each well, protected from light. Incubate for 15 to 20 minutes. Substrate Solution should remain colorless until added to the plate.

8.10 Quenching color development: Add 100 µL of Stop Solution to each well in the same order as addition of the TMB substrate. Mix by tapping the side of the plate gently. NB: Avoid skin and eye contact with the Stop solution.

8.11 Read results: Immediately after adding Stop solution read the absorbance on a microplate reader at a wavelength of 450 nm. If possible, perform a double wavelength readout (450 nm and 630 nm).

8.12 Data analysis: Calculate the average of the duplicate readings (OD value) for each standard and sample, and subtract the average of the zero standard absorbance. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis, use four-parameter logistic curve- fit (4-PL) analysis to do this. If the samples have been diluted, the OD readout from the standard curve must be multiplied by the dilution factor used.

Step	Reagent	Volume	Incubation	Wash	Notes
1	Standard and Samples	100 µL	120 min	4 times	Cover Wells incubate at 37°C
2	Diluent Antibody Solution	100 µL	60 min	4 times	Cover Wells incubate at 37°C
3	Diluent HRP Solution	100 µL	40 min	4 times	Cover Wells incubate at 37°C
4	TMB Substrate	100 µL	15-20 min	Do not wash	Incubate in the dark at 37°C
5	Stop Solution	100 µL	0 min	Do not wash	-
6	Read plate at 450 nm and 630 nm immediately after adding Stop solution. DO NOT exceed 5 minutes.				

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

9. Validation Data

9.1 Standard curve

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(ug/mL)	O.D	Average	Corrected

9.2 Precision

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested 20 times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays) Three samples of known concentration were tested in 24 separate assays to assess inter-assay precision.

Intra-assay Precision				
Sample	n	Mean (ug/mL)	SD	CV%
1	20	150.0	4.9	3.3
2	20	20.2	1.5	7.4

Inter-assay Precision				
Sample	n	Mean (ug/mL)	SD	CV%
1	24	149.7	15.9	10.6
2	24	19.9	2.1	10.6

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com

10. References

1. Wei, Jiangbin et al. "Transcriptome profiling of cells exposed to particular and intense electromagnetic radiation emitted by the "SG-III" prototype laser facility." Scientific reports vol. 11,1 2017. 21 Jan. 2021, doi:10.1038/s41598-021-81642-5

2. Zehbe, Ingeborg et al. "Rare human papillomavirus 16 E6 variants reveal significant oncogenic potential." Molecular cancer vol. 10 77. 24 Jun. 2011, doi:10.1186/1476-4598-10-77

3. Zhang, Liang et al. "Activated mitochondrial apoptosis in hESCs after dissociation involving the PKA/p-p53/Bax signaling pathway." Experimental cell research vol. 369,2 (2018): 226-233. doi:10.1016/j.yexcr.2018.05.024

4. Kang, Mi Ae et al. "A novel pyrido-thieno-pyrimidine derivative activates p53 through induction of phosphorylation and acetylation in colorectal cancer cells." *International journal of oncology vol. 46,1 (2015): 342-50. doi:10.3892/ijo.2014.2720*

USA: Proteintech@ptglab.com
UK & Europe: Europe@ptglab.com
China: Proteintech-CN@ptglab.com