

Human PAPP A2 Sandwich ELISA Kit Datasheet

Please read it entirely before use

Catalogue Number: KE01079

Size: 96T

Sensitivity: 0.01 ng/mL

Range: 0.156-10 ng/mL

Usage: For the quantitative detection of human PAPP A2 concentrations in serum, plasma and tissue lysate.


This product is for research use only and not for use in human or animal therapeutic or diagnostic.

Table of content	page
1. Background	3
2. Principle	3
3. Required Materials	3
4. Kit Components and Storage	4
5. Safety Notes	4
6. Sample Collection and Storage	4
7. Regent Preparation	5
8. Assay Procedure Summary	6
9. Validation Data	7
9.1 Standard curve	7
9.2 Precision	8
9.3 Recovery	8
9.4 Sample values	8
9.5 Sensitivity	8
9.6 Linearity	9
9.7 Specificity	9
10. References	9

1. Background

PAPP A2 (Pregnancy-associated Plasma Protein A2) is a metalloproteinase member of the PAPP A protein family and is mainly secreted by the syncytiotrophoblast cells of the placenta. Unlike PAPP A in the same family that regulates the biological activity of insulin-like growth factors (IGFs), PAPP A2 can specifically cleave and inactivate the IGF-binding protein 5 (IGFBP-5), thereby releasing more free IGFs to promote fetal development. Moreover, its concentration significantly increases in the maternal blood during pregnancy, making it a potential new biomarker for assessing placental function and fetal growth restriction.

2. Principle



Sandwich ELISA structure (Detection antibody labeled with HRP)

A capture antibody is pre-coated onto the bottom of wells which binds to analyte of interest. A detection antibody labeled with HRP also binds to the analyte. TMB acts as the HRP substrate and the solution color will change from colorless to blue. A stop solution containing sulfuric acid turns solution yellow. The color intensity is proportional to the quantity of bound protein which is measurable at 450 nm with the correction wavelength set at 630 nm.

3. Required Materials

- 3.1 A microplate reader capable of measuring absorbance at 450 nm with the correction wavelength set at 630 nm.
- 3.2 Calibrated, adjustable precision pipettes and disposable plastic tips. A manifold multi-channel pipette is recommended for large assays.
- 3.3 Plate washer: automated or manual.
- 3.4 Absorbent paper towels.
- 3.5 Glass or plastic tubes to prepare standard and sample dilutions.
- 3.6 Beakers and graduated cylinders.
- 3.7 Log-log or semi-log graph paper or computer and software for ELISA data analysis. A four-parameter logistic (4-PL) curve-fit is recommended.

4. Kit Components and Storage

Microplate - antibody coated 96-well microplate (8 well × 12 strips)	1 plate	Unopened Kit: Store at 2-8°C for 6 months or -20°C for 12 months. Opened Kit: All reagents stored at 2-8°C for 7 days. Please use a new standard for each assay.
Protein standard - 20 ng/bottle; lyophilized	2 bottles	
Detection antibody, HRP-conjugated (100×) - 120 µL/vial*	1 vial	
Sample Diluent PT 3 - 30 mL/bottle. For human serum and plasma.	1 bottle	
Sample Diluent PT 4B1 - 30 mL/bottle. For tissue lysate.	1 bottle	
Detection Diluent - 30 mL/bottle	1 bottle	
Wash Buffer Concentrate (20×) - 30 mL/bottle	1 bottle	
Extraction Reagent - 30 mL/bottle	1 bottle	
Tetramethylbenzidine Substrate (TMB) - 12 mL/bottle	1 bottle	
Stop Solution - 12 mL/bottle	1 bottle	
Plate Cover Seals	4 pieces	

* Centrifugation immediately before use

5. Safety Notes

- 5.1 Avoid any skin and eye contact with Stop Solution and TMB. In case of contact, wash thoroughly with water.
- 5.2 Do not use the kit after the expiration date.
- 5.3 Do not mix or substitute reagents or materials from other kit lots or other sources.
- 5.4 Be sure to wear protective equipment such as gloves, masks and goggles during the experiment.
- 5.5 When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer to improve assay precision

6. Sample Collection and Storage

- 6.1 Serum: Allow blood samples to clot for 30 minutes, followed by centrifugation for 15 minutes at 1000xg. Clear serum can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.
- 6.2 Plasma: Use EDTA, heparin, or citrate as an anticoagulant for plasma collection. Centrifuge for 15 minutes at 1000xg within 30 minutes of collection. The plasma can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.
- 6.3 Tissue Lysate:
 - 1) Rinse tissue with PBS, cut into 1-2 mm pieces.
 - 2) Add protease inhibitor cocktail to the Extraction Reagent to a final concentration immediately prior to performing tissue lysis.
 - 3) Add 1 mL of Extraction Reagent containing protease inhibitor cocktail per 100 mg tissue.
 - 4) Homogenize the tissue completely using desired method on ice, Incubate on ice for 30 minutes, use ultrasound to break up the cells.
 - 5) Centrifuge tissue homogenates at 10,000 x g for 5 minutes at 4°C. Collect the supernatant, assay immediately or aliquot and store at -20°C.
 - 6) Measure the concentration of total protein in tissue homogenates using BCA assay.
 - 7) Avoid protein degradation by performing all the above procedures on ice where possible.

7. Regent Preparation

7.1 Wash Buffer (1X): If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 30 mL of Wash Buffer Concentrate(20X) to 570 mL deionized or distilled water to prepare 1X Wash Buffer.


7.2 Detection Antibody, HRP-conjugated(1X): Dilute 100X Detection Antibody, HRP-conjugated 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 μ L 100X Detection Antibody, HRP-conjugated + 990 μ L Detection Diluent (Centrifuge the 100X Detection Antibody solution, HRP-conjugated for a few seconds prior to use).

7.3 Sample Dilution: Different samples should be diluted with corresponding Sample Diluent, samples may require further dilution if the readout values are higher than the highest standard OD reading. Variations in sample collection, processing and storage may affect the results of the measurement.

Recommended Dilution for different sample types: 1:2 is recommended for human serum and plasma; 1:32 or 1:64 is recommended for tissue lysate.

7.4 Standard Serial Dilution:

For human serum and plasma, add 2 mL Sample Diluent PT 3 in protein standard. For tissue lysate, add 2 mL Sample Diluent PT 4B1 in protein standard.



Add # μ L of Standard diluted in the previous step	—	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
# μ L of Sample Diluent PT 3 or PT 4B1	2000 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
	"sd7"	"sd6"	"sd5"	"sd4"	"sd3"	"sd2"	"sd1"

8. Assay Procedure Summary

Bring all reagents to room temperature before use (Detection antibody, HRP-conjugated can be used immediately). To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

8.1 Take out the required number of microplate strips and return excess strips to the foil pouch containing the drying reagent pack and reseal; store at 4°C immediately. Microplate strips should be used in one week.

8.2 Preset the layout of the microplate, including control group, standard group and sample group, add 100 µL of each standard and sample to the appropriate wells. (Make sure sample addition is uninterrupted and completed within 5 to 10 minutes, it is recommended to assay all standards, controls, and samples in duplicate).

8.3 Seal plate with cover seal, pressing it firmly onto top of microwells. Incubate the plate for 2 hours at 37°C.

8.4 Wash

1) Gently remove the cover seal. Discard the liquid from wells by aspirating or decanting. Remove any residual solution by tapping the plate a few times on fresh paper towels.

2) Wash 4 times with 1X Wash Buffer, using at least 350-400 µL per well. Following the last wash, firmly tap plates on fresh towels 10 times to remove residual Wash Buffer. Avoid getting any towel fibers in the wells or wells drying out completely.

8.5 Add 100 µL of 1X Detection antibody, HRP-conjugated solution (refer to Reagent Preparation 7.2) to each well. Seal plate with cover seal and incubate for 40 minutes at 37°C.

8.6 Repeat wash step in 8.4.

8.7 Signal development: Add 100 µL of TMB substrate solution to each well, protected from light. Incubate for 15 to 20 minutes. Substrate Solution should remain colorless until added to the plate.

8.8 Quenching color development: Add 100 µL of Stop Solution to each well in the same order as addition of the TMB substrate. Mix by tapping the side of the plate gently. NB: Avoid skin and eye contact with the Stop solution.

8.9 Read results: Immediately after adding Stop solution read the absorbance on a microplate reader at a wavelength of 450 nm. If possible, perform a double wavelength readout (450 nm and 630 nm).


8.10 Data analysis: Calculate the average of the duplicate readings (OD value) for each standard and sample, and subtract the average of the zero standard absorbance. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis, use four-parameter logistic curve-fit (4-PL) analysis to do this. If the samples have been diluted, the OD readout from the standard curve must be multiplied by the dilution factor used.

Step	Reagent	Volume	Incubation	Wash	Notes
1	Standard and Samples	100 µL	120 min	4 times	Cover Wells incubate at 37°C
2	Diluent Detection antibody, HRP-conjugated Solution	100 µL	40 min	4 times	Cover Wells incubate at 37°C
3	TMB Substrate	100 µL	15-20 min	Do not wash	Incubate in the dark at 37°C
4	Stop Solution	100 µL	0 min	Do not wash	-
5	Read plate at 450 nm and 630 nm immediately after adding Stop solution. DO NOT exceed 5 minutes.				


9. Validation Data

9.1 Standard curve

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(ng/mL)	O.D	Average	Corrected
0	0.0154 0.0147	0.0151	-
0.156	0.0477 0.0383	0.0430	0.0280
0.313	0.0778 0.0814	0.0796	0.0646
0.625	0.1845 0.1679	0.1762	0.1612
1.25	0.3422 0.3217	0.3320	0.3169
2.5	0.6933 0.6306	0.6620	0.6469
5	1.2288 1.1632	1.1960	1.1810
10	1.9856 2.0344	2.0100	1.9950



(ng/mL)	O.D	Average	Corrected
0	0.0095 0.0085	0.0090	-
0.156	0.0433 0.0467	0.0450	0.0360
0.313	0.0999 0.0975	0.0987	0.0897
0.625	0.201 0.2038	0.2024	0.1934
1.25	0.3831 0.3987	0.3909	0.3819
2.5	0.772 0.7456	0.7588	0.7498
5	1.3592 1.3813	1.3703	1.3613
10	2.2 2.234	2.2170	2.2080

9.2 Precision

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested 8 times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays) Three samples of known concentration were tested in 16 separate assays to assess inter-assay precision.

Intra-assay Precision					Inter-assay Precision				
Sample	n	Mean (ng/mL)	SD	CV%	Sample	n	Mean (ng/mL)	SD	CV%
1	8	5.39	0.11	2.04	1	16	5.12	0.38	7.42
2	8	1.32	0.07	5.30	2	16	1.30	0.09	6.92
3	8	0.61	0.04	6.56	3	16	0.63	0.05	7.94

9.3 Recovery

The recovery of human PAPP A2 spiked to three different levels throughout the range of the assay in various matrices was evaluated.

Sample Type		Average% of Expected	Range (%)
Human plasma	1:2	125	121-127
Tissue lysate	1:256	99	94-103

9.4 Sample values

Human plasma - human plasma samples were evaluated for the presence of human PAPP A2 in this assay.

Sample Type	Mean (ng/mL)	% Detectable	Range (ng/mL)
Human plasma (n=16)	0.49	31	ND-1.94

ND*=Non-detectable

Tissue lysate

	Human PAPP A2 (ng/mL)	Total protein (mg/mL)
Human placenta tissue lysate	263.87	1.00

9.5 Sensitivity

The minimum detectable dose of human PAPP A2 is 0.01 ng/mL. This was determined by adding two standard deviations to the concentration corresponding to the mean O.D. of 20 zero standard replicates.

9.6 Linearity

To assess the linearity of the assay, human plasma samples were spiked with high concentrations of human PAPP A2 and diluted with the appropriate **Sample Diluent** to produce samples with values within the dynamic range of the assay. Tissue lysate samples were diluted with the appropriate **Sample Diluent** to produce samples with values within the dynamic range of the assay.

(The tissue lysate was initially diluted 1:16.)

		Human plasma (Sample Diluent PT 3)	Tissue lysate (Sample Diluent PT 4B1)
1:2	Average% of Expected	92	100
	Range (%)	76-108	-
1:4	Average% of Expected	89	113
	Range (%)	75-103	112-113
1:8	Average% of Expected	89	117
	Range (%)	78-99	114-119
1:16	Average% of Expected	84	129
	Range (%)	77-90	122-136

9.7 Specificity

This kit specifically recognizes native and recombinant human PAPP A2.

10. References

1. Barrios, Vicente et al. Cells vol. 10,12 3576. 18 Dec. 2021,
2. Nimptsch, Katharina et al. Scientific reports vol. 14,1 1770. 20 Jan. 2024