

Mouse/Rat IGF1 Sandwich ELISA Kit Datasheet

Please read it entirely before use

Catalogue Number: KE20037

Size: 96T

Sensitivity: 0.2 pg/mL

Range: 1.95-125 pg/mL

Usage: For the quantitative detection of mouse/rat IGF1 concentrations in serum, plasma, cell culture supernatant and tissue lysate.


This product is for research use only and not for use in human or animal therapeutic or diagnostic.

Table of content	page
1. Background	3
2. Principle	3
3. Required Materials	3
4. Kit Components and Storage	4
5. Safety Notes	4
6. Sample Collection and Storage	5
7. Regent Preparation	6
8. Assay Procedure Summary	7
9. Validation Data	8
9.1 Standard curve	8
9.2 Precision	8
9.3 Recovery	9
9.4 Sample values	9
9.5 Sensitivity	9
9.6 Linearity	10
9.7 Specificity	10
10. References	10

1. Background

Insulin-like-growth factor 1 (IGF1), a 70 amino-acid peptide hormone, is the principal mediator of biochemical effects of growth hormone (GH) and important in the regulation of cell proliferation and differentiation. IGF1 is largely synthesized in the liver (75%) and, to a lesser extent, in peripheral tissues. IGF1 is a potent mitogen and is inhibited by IGF-binding protein-3 (IGFBP3). High serum IGF1 and low IGFBP3 are associated with increased risk of several carcinomas. Mice lacking IGF1 exhibit generalized organ hypoplasia including underdevelopment of the central nervous system and developmental defects in bone, muscle and reproductive systems.

2. Principle



Sandwich ELISA structure (Detection antibody labeled with HRP)

A capture antibody is pre-coated onto the bottom of wells which binds to analyte of interest. A detection antibody labeled with HRP also binds to the analyte. TMB acts as the HRP substrate and the solution color will change from colorless to blue. A stop solution containing sulfuric acid turns solution yellow. The color intensity is proportional to the quantity of bound protein which is measurable at 450 nm with the correction wavelength set at 630 nm.

3. Required Materials

- 3.1 A microplate reader capable of measuring absorbance at 450 nm with the correction wavelength set at 630 nm.
- 3.2 Calibrated, adjustable precision pipettes and disposable plastic tips. A manifold multi-channel pipette is recommended for large assays.
- 3.3 Plate washer: automated or manual.
- 3.4 Absorbent paper towels.
- 3.5 Glass or plastic tubes to prepare standard and sample dilutions.
- 3.6 Beakers and graduated cylinders.
- 3.7 Log-log or semi-log graph paper or computer and software for ELISA data analysis. A four-parameter logistic (4-PL) curve-fit is recommended.

4. Kit Components and Storage

Microplate - antibody coated 96-well microplate (8 well × 12 strips)	1 plate	Unopened Kit: Store at 2-8°C for 6 months or -20°C for 12 months. Opened Kit: All reagents stored at 2-8°C for 7 days. Please use a new standard for each assay.
Protein standard - 250 pg/bottle; lyophilized	2 bottles	
Detection antibody, HRP-conjugated (100×) - 120 µL/vial*	1 vial	
Sample Diluent PT 4B1 - 30 mL/bottle	2 bottles	
Detection Diluent - 30 mL/bottle	1 bottle	
Wash Buffer Concentrate (20×) - 30 mL/bottle	1 bottle	
Extraction Reagent - 30 mL/bottle	1 bottle	
Tetramethylbenzidine Substrate (TMB) - 12 mL/bottle	1 bottle	
Stop Solution - 12 mL/bottle	1 bottle	
Plate Cover Seals	4 pieces	

* Centrifugation immediately before use

5. Safety Notes

5.1 Avoid any skin and eye contact with Stop Solution and TMB. In case of contact, wash thoroughly with water.

5.2 Do not use the kit after the expiration date.

5.3 Do not mix or substitute reagents or materials from other kit lots or other sources.

5.4 Be sure to wear protective equipment such as gloves, masks and goggles during the experiment.

5.5 When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer to improve assay precision

6. Sample Collection and Storage

6.1 Serum: Allow blood samples to clot for 30 minutes, followed by centrifugation for 15 minutes at 1000xg. Clear serum can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.

6.2 Plasma: Use EDTA, heparin, or citrate as an anticoagulant for plasma collection. Centrifuge for 15 minutes at 1000xg within 30 minutes of collection. The plasma can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.

6.3 Cell Culture Supernatant: Remove particulates by centrifugation for 5 minutes at 500xg and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

6.4 Tissue Lysate:

- 1) Rinse tissue with PBS, cut into 1-2 mm pieces.
- 2) Add protease inhibitor cocktail to the Extraction Reagent to a final concentration immediately prior to performing tissue lysis.
- 3) Add 1 mL of Extraction Reagent containing protease inhibitor cocktail per 100 mg tissue.
- 4) Homogenize the tissue completely using desired method on ice, Incubate on ice for 30 minutes, use ultrasound to break up the cells.
- 5) Centrifuge tissue homogenates at 10,000 x g for 5 minutes at 4°C. Collect the supernatant, assay immediately or aliquot and store at -20°C.
- 6) Measure the concentration of total protein in tissue homogenates using BCA assay.
- 7) Avoid protein degradation by performing all the above procedures on ice where possible.

7. Regent Preparation

7.1 Wash Buffer (1X): If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 30 mL of Wash Buffer Concentrate(20X) to 570 mL deionized or distilled water to prepare 1X Wash Buffer.


7.2 Detection Antibody, HRP-conjugated(1X): Dilute 100X Detection Antibody, HRP-conjugated 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 μ L 100X Detection Antibody, HRP-conjugated + 990 μ L Detection Diluent (Centrifuge the 100X Detection Antibody solution, HRP-conjugated for a few seconds prior to use).

7.3 Sample Dilution: Different samples should be diluted with corresponding Sample Diluent, samples may require further dilution if the readout values are higher than the highest standard OD reading. Variations in sample collection, processing and storage may affect the results of the measurement.

Recommended Dilution for different sample types: 1:5,000 or 1:10,000 is recommended for mouse serum and plasma; 1:10,000 or 1:20,000 is recommended for rat serum and plasma; 1:64 or 1:128 is recommended for cell culture supernatant; 1:64 or 1:128 is recommended for tissue lysate.

7.4 Standard Serial Dilution:

Add 2 mL Sample Diluent PT 4B1 in protein standard.



Add # μ L of Standard diluted in the previous step	—	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
# μ L of Sample Diluent PT 4B1	2000 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
	"sd7"	"sd6"	"sd5"	"sd4"	"sd3"	"sd2"	"sd1"

8. Assay Procedure Summary

Bring all reagents to room temperature before use (Detection antibody, HRP-conjugated can be used immediately). To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

8.1 Take out the required number of microplate strips and return excess strips to the foil pouch containing the drying reagent pack and reseal; store at 4°C immediately. Microplate strips should be used in one week.

8.2 Preset the layout of the microplate, including control group, standard group and sample group, add 100 µL of each standard and sample to the appropriate wells. (Make sure sample addition is uninterrupted and completed within 5 to 10 minutes, It is recommended to assay all standards, controls, and samples in duplicate).

8.3 Seal plate with cover seal, pressing it firmly onto top of microwells. Incubate the plate for 2 hours at 37°C.

8.4 Wash

1) Gently remove the cover seal. Discard the liquid from wells by aspirating or decanting. Remove any residual solution by tapping the plate a few times on fresh paper towels.

2) Wash 4 times with 1X Wash Buffer, using at least 350-400 µL per well. Following the last wash, firmly tap plates on fresh towels 10 times to remove residual Wash Buffer. Avoid getting any towel fibers in the wells or wells drying out completely.

8.5 Add 100 µL of 1X Detection antibody, HRP-conjugated solution (refer to Reagent Preparation 7.2) to each well. Seal plate with cover seal and incubate for 40 minutes at 37°C.

8.6 Repeat wash step in 8.4.

8.7 Signal development: Add 100 µL of TMB substrate solution to each well, protected from light. Incubate for 15 to 20 minutes. Substrate Solution should remain colorless until added to the plate.

8.8 Quenching color development: Add 100 µL of Stop Solution to each well in the same order as addition of the TMB substrate. Mix by tapping the side of the plate gently. NB: Avoid skin and eye contact with the Stop solution.

8.9 Read results: Immediately after adding Stop solution read the absorbance on a microplate reader at a wavelength of 450 nm. If possible, perform a double wavelength readout (450 nm and 630 nm).


8.10 Data analysis: Calculate the average of the duplicate readings (OD value) for each standard and sample, and subtract the average of the zero standard absorbance. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis, use four-parameter logistic curve-fit (4-PL) analysis to do this. If the samples have been diluted, the OD readout from the standard curve must be multiplied by the dilution factor used.

Step	Reagent	Volume	Incubation	Wash	Notes
1	Standard and Samples	100 µL	120 min	4 times	Cover Wells incubate at 37°C
2	Diluent Detection antibody, HRP-conjugated Solution	100 µL	40 min	4 times	Cover Wells incubate at 37°C
3	TMB Substrate	100 µL	15-20 min	Do not wash	Incubate in the dark at 37°C
4	Stop Solution	100 µL	0 min	Do not wash	-
5	Read plate at 450 nm and 630 nm immediately after adding Stop solution. DO NOT exceed 5 minutes.				

9. Validation Data

9.1 Standard curve

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(pg/mL)	O.D	Average	Corrected
0	0.027 0.028	0.028	-
1.95	0.1117 0.1214	0.117	0.089
3.9	0.1908 0.2178	0.204	0.177
7.8	0.3533 0.3771	0.365	0.338
15.6	0.6846 0.6997	0.692	0.665
31.25	1.106 1.1511	1.129	1.101
62.5	1.8585 1.8788	1.869	1.841
125	2.5044 2.4962	2.5	2.473

9.2 Precision

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested 8 times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays) Three samples of known concentration were tested in 16 separate assays to assess inter-assay precision.

Intra-assay Precision				
Sample	n	Mean (pg/mL)	SD	CV%
1	8	60.8	2.2	3.6
2	8	16.0	0.4	2.4
3	8	8.2	0.2	2.0

Inter-assay Precision				
Sample	n	Mean (pg/mL)	SD	CV%
1	16	62.6	2.8	4.5
2	16	16.5	0.7	4.0
3	16	8.4	0.3	3.8

9.3 Recovery

The recovery of mouse/rat IGF1 spiked to three different levels throughout the range of the assay in various matrices was evaluated.

Sample Type		Average% of Expected	Range (%)
Mouse serum	1:20,000	102	97-106
	1:40,000	97	93-99
Rat serum	1:80,000	89	87-91
Cell culture supernatant	1:256	92	86-95
	1:512	104	97-117
Mouse tissue lysate	1:256	95	81-105
	1:512	94	86-100
Rat tissue lysate	1:256	98	95-101

9.4 Sample values

Mouse/Rat serum - mouse and rat serum samples were evaluated for the presence of IGF1 in this assay.

Sample Type	Mean (ng/mL)	Range (ng/mL)
Mouse serum (n=16)	201.8	141.1-294.1
Rat serum (n=16)	1,108.1	649.3-1,499.5

Cell culture supernatant - rat heart, cut into 1-2 mm pieces, was cultured for 4 days in 30 mL of RPMI supplemented with 10% fetal bovine serum and stimulated with 2.5 ng/mL LPS. Aliquots of the cell culture supernates were removed and assayed for IGF1, and measured 3,623.3 pg/mL.

Tissue lysate

	IGF1 (pg/mL)	Total protein (mg/mL)
Mouse lung tissue lysate	2,676.7	5.3
Mouse heart tissue lysate	5,875.6	7.1
Rat lung tissue lysate	2,676.0	1.2
Rat heart tissue lysate	5,101.2	6.8

9.5 Sensitivity

The minimum detectable dose of mouse/rat IGF1 is 0.2 pg/mL. This was determined by adding two standard deviations to the concentration corresponding to the mean O.D. of 20 zero standard replicates.

9.6 Linearity

To assess the linearity of the assay, samples were diluted with the appropriate **Sample Diluent** to produce samples with values within the dynamic range of the assay.

(The mouse serum was initially diluted 1:2,500. The rat serum was initially diluted 1:5,000. The cell culture supernatant was initially diluted 1:32. The mouse tissue lysate was initially diluted 1:32. The rat tissue lysate was initially diluted 1:32.)

		Mouse serum	Rat serum	Cell culture supernatant	Mouse tissue lysate	Rat tissue lysate
1:2	Average% of Expected	100	100	100	100	100
	Range (%)	-	-	-	-	-
1:4	Average% of Expected	104	107	101	100	103
	Range (%)	103-105	105-109	100-102	99-100	97-108
1:8	Average% of Expected	107	110	99	109	113
	Range (%)	99-115	106-113	98-99	108-110	106-119
1:16	Average% of Expected	113	112	95	119	113
	Range (%)	110-116	111-113	94-96	117-120	105-121

9.7 Specificity

This assay recognizes natural and recombinant mouse/rat IGF1.

The following factors prepared at 50 ng/mL were assayed and exhibited no cross-reactivity or interference.

Recombinant human:

IGFBP-2

IGFBP-3

IGFBP-4

Recombinant mouse:

EGF

EGFR

IGFBP-1

IGFBP-2

IGFBP-3

10. References

1. Adamo, M L et al. Advances in experimental medicine and biology vol. 343 (1993): 1-11.
2. Adachi, Yasushi et al. Journal of gastroenterology and hepatology vol. 34,12 (2019): 2104-2111.
3. Anisimov, Vladimir N, and Andrzej Bartke. Critical reviews in oncology/hematology vol. 87,3 (2013): 201-23.
4. Sonntag, William E et al. The journals of gerontology. Series A, Biological sciences and medical sciences vol. 67,6 (2012): 587-98.