

Mouse TNFRSF1B Sandwich ELISA Kit Datasheet

Please read it entirely before use

Catalogue Number: KE10061 Size: 96T Sensitivity: 0.1 pg/mL Range: 3.9-250 pg/mL Usage: For the quantitative detection of mouse TNFRSF1B concentrations in serum, plasma and cell culture supernatant.

This product is for research use only and not for use in human or animal therapeutic or diagnostic.




Table of content	pa	ge
1. Background		• 3
2. Principle		• 3
3. Required Materials		• 3
4. Kit Components and Storage		4
5. Safety Notes		• 4
6. Sample Collection and Storage		• 4
7. Regent Preparation		• 5
8. Assay Procedure Summary		• 6
9. Validation Data		• 7
9.1 Standard curve		• 7
9.2 Precision		• 7
9.3 Recovery		• 8
9.4 Sample values		8
9.5 Sensitivity		• 8
9.6 Linearity		• 9
10. References		• 9

ptglab.com

1. Background

TNFRSF1B, also known as TNFR2 or CD120b or p75, is a member of the tumor necrosis factor receptor superfamily, which also contains TNFRSF1A. Tumor necrosis factor (TNF, TNFSF2) is a multifunctional cytokine that plays a key role in regulating inflammation, immune functions, host defense, and apoptosis. TNF signals through TNFRSF1A and TNFRSF1B. TNFRSF1A is widely expressed, whereas TNFRSF1B exhibits more restricted expression, being found on CD4 and CD8 T lymphocytes, endothelial cells, microglia, oligodendrocytes, neuron subtypes, cardiac myocytes, thymocytes and human mesenchymal stem cells. Both TNFRSF1A and TNFRSF1B can also exist in soluble forms, probably derived by proteolytic cleavage from the cell surface forms. Various defects in the TNFRSF1B pathway, due to polymorphisms in the TNFRSF1B gene, upregulated expression of TNFRSF1B and TNFRSF1B shedding, have been implicated in the pathology of several autoimmune disorders.

2. Principle



Sandwich ELISA structure (Detection antibody labeled with biotin)

A capture antibody is pre-coated onto the bottom of wells which binds to analyte of interest. A detection antibody labeled with biotin also binds to the analyte. Streptavidin-HRP binds to the biotin. TMB acts as the HRP substrate and the solution color will change from colorless to blue. A stop solution containing sulfuric acid turns solution yellow. The color intensity is proportional to the quantity of bound protein which is measurable at 450 nm with the correction wavelength set at 630 nm.

3. Required Materials

3.1 A microplate reader capable of measuring absorbance at 450 nm with the correction wavelength set at 630 nm.

3.2 Calibrated, adjustable precision pipettes and disposable plastic tips. A manifold multi-channel pipette is recommended for large assays.

- 3.3 Distilled or deionized water.
- 3.4 Plate washer: automated or manual.
- 3.5 Absorbent paper towels.
- 3.6 Glass or plastic tubes to prepare standard and sample dilutions.

3.7 Beakers and graduated cylinders.

3.8 100 mM PMSF (phenylmethanesulfonyl fluoride) stock solution (1g PMSF in 57 mL isopropyl alcohol), only for lysate sample. 3.9 Log-log or semi-log graph paper or computer and software for ELISA data analysis. A four-parameter logistic (4-PL) curve-fit is recommended.

ptglab.com

4. Kit Components and Storage

Microplate - antibody coated 96-well microplate (8 well × 12 strips)	1 plate	Unopened Kit:
Protein standard - 250 pg/bottle; lyophilized	2 bottles	•
Detection antibody, biotinylated (100×) - 120 µL/vial*	1 vial	Store at 2-8°C for 6 months or -
Streptavidin-horseradish peroxidase (HRP) (100×) - 120 µL/vial*	1 vial	20°C for 12 months.
Sample Diluent PT 4 - 30 mL/bottle.	2 bottles	Opened Kit:
Detection Diluent - 30 mL/bottle	1 bottle	All reagents stored at 2-8°C for
Wash Buffer Concentrate (20×) - 30 mL/bottle	1 bottle	C C
Tetramethylbenzidine Substrate (TMB) - 12 mL/bottle	1 bottle	7 days.
Stop Solution - 12 mL/bottle	1 bottle	Please use a new standard
Plate Cover Seals	4 pieces	for each assay.

* Centrifugation immediately before use

5. Safety Notes

5.1 Avoid any skin and eye contact with Stop Solution and TMB. In case of contact, wash thoroughly with water.

5.2 Do not use the kit after the expiration date.

5.3 Do not mix or substitute reagents or materials from other kit lots or other sources.

5.4 Be sure to wear protective equipment such as gloves, masks and goggles during the experiment.

5.5 When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer to improve assay precision

6. Sample Collection and Storage

6.1 Serum: Allow blood samples to clot for 30 minutes, followed by centrifugation for 15 minutes at 1000xg. Clear serum can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.

6.2 Plasma: Use EDTA, heparin, or citrate as an anticoagulant for plasma collection. Centrifuge for 15 minutes at 1000xg within 30 minutes of collection. The plasma can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.

6.3 Cell Culture Supernatant: Remove particulates by centrifugation for 5 minutes at 500xg and assay immediately or aliquot and store samples at \leq -20°C. Avoid repeated freeze-thaw cycles.

7. Regent Preparation

7.1 Wash Buffer (1X): If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 30 mL of Wash Buffer Concentrate(20X) to 570 mL deionized or distilled water to prepare 1X Wash Buffer.

7.2 Detection Antibody (1X): Dilute 100X Detection Antibody 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 µL 100X Detection Antibody + 990 µL Detection Diluent (Centrifuge the 100 X Detection Antibody solution for a few seconds prior to use).


7.3 Streptavidin-HRP (1X): Dilute 100X Streptavidin-HRP 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 µL 100X Streptavidin-HRP + 990 µL Detection Diluent (Centrifuge the 100X Streptavidin-HRP solution for a few seconds prior to use).

7.4 Sample Dilution: Different samples should be diluted with corresponding Sample Diluent, samples may require further dilution if the readout values are higher than the highest standard OD reading. Variations in sample collection, processing and storage may affect the results of the measurement.

Recommended Dilution for different sample types: 1:80 or 1:160 is recommended for mouse serum and plasma; 1:8 or 1:16 is recommended for cell culture supernatant.

7.5 Standard Serial Dilution:

Add 1 mL Sample Diluent PT 4 in protein standard.



Add # μL of Standard diluted in the previous step	-	500 µL	500 μL	500 μL	500 μL	500 μL	500 µL
#μL of Sample Diluent PT 4	1000 µL	500 µL	500 μL	500 μL	500 μL	500 μL	500 µL
	"sd7"	"sd6"	"sd5"	"sd4"	"sd3"	"sd2"	"sd1"

8. Assay Procedure Summary

Bring all reagents to room temperature before use (Detection antibody and Streptavidin-HRP can be used immediately). To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent. 8.1 Take out the required number of microplate strips and return excess strips to the foil pouch containing the drying reagent pack and reseal; store at 4°C immediately. Microplate strips should be used in one week.

8.2 Preset the layout of the microplate, including control group, standard group and sample group, add 100 µL of each standard and sample to the appropriate wells. (Make sure sample addition is uninterrupted and completed within 5 to 10 minutes, It is recommended to assay all standards, controls, and samples in duplicate).

8.3 Seal plate with cover seal, pressing it firmly onto top of microwells. Incubate the plate for 2 hours at 37°C.8.4 Wash

1) Gently remove the cover seal. Discard the liquid from wells by aspirating or decanting. Remove any residual solution by tapping the plate a few times on fresh paper towels.

2) Wash 4 times with 1X Wash Buffer, using at least 350-400 µL per well. Following the last wash, firmly tap plates on fresh towels 10 times to remove residual Wash Buffer. Avoid getting any towel fibers in the wells or wells drying out completely.
8.5 Add 100 µL of 1X Detection Antibody solution (refer to Reagent Preparation7.2) to each well. Seal plate with cover seal and incubate for 1 hour at 37°C.

8.6 Repeat wash step in 8.4.

8.7 Add 100 μ L of 1X Streptavidin-HRP solution (refer to Reagent Preparation7.3) to each well. Seal plate with cover seal and incubate the plate for 40 minutes at 37°C.

8.8 Repeat wash step in 8.4.

8.9 Signal development: Add 100 μL of TMB substrate solution to each well, protected from light. Incubate for 15 to 20 minutes. Substrate Solution should remain colorless until added to the plate.

8.10 Quenching color development: Add 100 µL of Stop Solution to each well in the same order as addition of the TMB substrate. Mix by tapping the side of the plate gently. NB: Avoid skin and eye contact with the Stop solution.

8.11 Read results: Immediately after adding Stop solution read the absorbance on a microplate reader at a wavelength of 450 nm. If possible, perform a double wavelength readout (450 nm and 630 nm).


8.12 Data analysis: Calculate the average of the duplicate readings (OD value) for each standard and sample, and subtract the average of the zero standard absorbance. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis, use four-parameter logistic curve- fit (4-PL) analysis to do this. If the samples have been diluted, the OD readout from the standard curve must be multiplied by the dilution factor used.

Step	Reagent	Volume	Incubation	Wash	Notes	
1	Standard and Samples	100 µL	120 min	4 times	Cover Wells incubate at 37°C	
2	Diluent Antibody Solution	100 µL	60 min	4 times	Cover Wells incubate at 37°C	
3	Diluent HRP Solution	100 µL	40 min	4 times	Cover Wells incubate at 37°C	
4	TMB Substrate	100 µL	15-20 min	Do not wash	Incubate in the dark at 37°C	
5	Stop Solution	100 µL	0 min	Do not wash	-	
6	6 Read plate at 450 nm and 630 nm immediately after adding Stop solution. DO NOT exceed 5 minutes.					

9. Validation Data

9.1 Standard curve

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(pg/mL)	0.D	Average	Corrected
0	0.028 0.026	0.027	-
3.9	0.096 0.109	0.103	0.076
7.8	0.184 0.188	0.186	0.159
15.6	0.342 0.347	0.345	0.318
31.25	0.645 0.671	0.658	0.631
62.5	1.179 1.177	1.178	1.151
125	1.897 1.888	1.893	1.866
250	2.510 2.603	2.557	2.53

9.2 Precision

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested 20 times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays) Three samples of known concentration were tested in 24 separate assays to assess inter-assay precision.

Intra-assay Precision						Inter-assay Precision				
Sample	n	Mean (pg/mL)	SD	CV%		Sample	n	Mean (pg/mL)	SD	CV%
1	20	108.5	3.9	3.6		1	24	92.9	8.6	9.3
2	20	33.3	0.7	2.0		2	24	30.7	2.4	7.9
3	20	15.6	0.5	3.2		3	24	14.5	1.3	9.3

9.3 Recovery

The recovery of mouse TNFRSF1B spiked to three different levels throughout the range of the assay in various matrices was evaluated.

Sample Type		Average% of Expected	Range (%)
Mouse corum	1:300	111	99-123
Mouse serum	1:600	107	103-112
Coll culture supernatant	1:30	120	111-128
Cell culture supernatant	1:60	115	109-122

9.4 Sample values

Mouse serum - The mouse serum samples were evaluated for the presence of mouse TNFRSF1B in this assay.

Sample Type	Mean of Detectable (pg/mL)	Range (pg/mL)
Mouse serum (n=16)	6,345.5	5,311.9-7,676.5

Cell culture supernatant - Mouse lung conditioned media cultured in RPMI supplemented with 10% fetal bovine serum was collected after culturing for 5 days. An aliquot of the cell culture supernatant was removed, assayed for mouse TNFRS1B and measured 596.6 pg/mL.

9.5 Sensitivity

The minimum detectable dose of mouse TNFRSF1B is 0.1 pg/mL. This was determined by adding two standard deviations to the concentration corresponding to the mean O.D. of 20 zero standard replicates.



9.6 Linearity

To assess the linearity of the assay, samples were diluted with the appropriate **Sample Diluent** to produce samples with values within the dynamic range of the assay.

(The mouse serum samples were initially diluted 1:40. The cell culture supernatant was initially diluted 1:4)

		Mouse serum	Cell culture supernatant
1.2	Average% of Expected	100	100
1:2	Range (%)	-	-
1./	Average% of Expected	104	102
1:4	Range (%)	99-107	98-107
1:8	Average% of Expected	104	102
1.0	Range (%)	101-110	95-112
1:16	Average% of Expected	104	98
1.10	Range (%)	99-107	93-106

10. References

1. Islam A, Adamik B, Hawari FI, et al. Extracellular TNFR1 release requires the calcium-dependent formation of a nucleobindin 2-ARTS-1 complex. J Biol Chem. 2006;281(10):6860-6873.

2. Faustman D, Davis M. TNF receptor 2 pathway: drug target for autoimmune diseases. Nat Rev Drug Discov. 2010;9(6):482-493.

3. Cabal-Hierro L, Lazo PS. Signal transduction by tumor necrosis factor receptors. Cell Signal. 2012;24(6):1297-1305.

4. Diez-Ruiz A, Tilz GP, Zangerle R, Baier-Bitterlich G, Wachter H, Fuchs D. Soluble receptors for tumour necrosis factor in clinical laboratory diagnosis. Eur J Haematol. 1995;54(1):1-8.

